

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
35 °C	35.357 °C	35.6 °C	0.243 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.202 %HR	32.8 %HR	2.598 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	52.7 %HR	2.672 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.568 %HR	70.4 %HR	0.832 %HR	± 3.0 %HR	± 1.9E+00 %HR