

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-5 °C	-5.295 °C	-5.0 °C	0.295 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.329 °C	25.5 °C	0.171 °C	± 0.5 °C	± 4.0E-01 °C
50 °C	50.037 °C	50.1 °C	0.063 °C	± 0.5 °C	± 1.1E+00 °C
20 %HR	19.977 %HR	22.8 %HR	2.823 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	52.7 %HR	2.672 %HR	± 3.0 %HR	± 1.9E+00 %HR
90 %HR	90.018 %HR	88.0 %HR	-2.018 %HR	± 3.0 %HR	± 1.9E+00 %HR