

Patrón (Standard)	UBP (UT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
3600.92 s	3600 s	-0.92 s	-0.026%	± 0.01 %	± 1.0E+03 s
3600.92 s	3600 s	-0.92 s	-0.026%	± 0.01 %	± 1.0E+03 s
3600.92 s	3600 s	-0.92 s	-0.026%	± 0.01 %	± 1.0E+03 s