

Patrón (Standard)	UBP (UT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
1800.174 s	1800 s	-0.174 s	-0.010%	± 0.01 % ± 1.0E+03 s
2700.057 s	2700 s	-0.057 s	-0.002%	± 0.01 % ± 1.0E+03 s
3599.992 s	3600 s	0.008 s	0.000%	± 0.01 % ± 1.0E+03 s