

<b>Patrón (Standard)</b>	<b>UBP (UT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
1799.778 s	1800 s	0.222 s	0.012%	± 0.01 %	± 1.0E+03 s
2459.874 s	2460 s	0.126 s	0.005%	± 0.01 %	± 1.0E+03 s
3599.858 s	3600 s	0.142 s	0.004%	± 0.01 %	± 1.0E+03 s