

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.208 °C	10.2 °C	-0.008 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.450 °C	25.4 °C	-0.050 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.564 °C	30.4 °C	-0.164 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.980 °C	40.0 °C	0.020 °C	± 0.5 °C	± 4.0E-01 °C

