

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-30 °C	-30.334 °C	-30.5 °C	-0.166 °C	± 0.5 °C	± 4.0E-01 °C
-20 °C	-20.276 °C	-20.0 °C	0.276 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.048 °C	-10.2 °C	-0.152 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	0.040 °C	-0.2 °C	-0.240 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.037 °C	10.1 °C	0.063 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.0 °C	-0.249 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.076 °C	35.0 °C	-0.076 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.284 %HR	41.40 %HR	1.116 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	55.417 %HR	57.60 %HR	2.183 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.049 %HR	76.30 %HR	1.251 %HR	± 3.0 %HR	± 1.9E+00 %HR