

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>
1.00 mm	1.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.50 mm	1.50 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
3.00 mm	2.99 mm	-10 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
5.00 mm	4.99 mm	-10 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
8.00 mm	7.99 mm	-10 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
10.00 mm	9.98 mm	-20 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
15.00 mm	14.98 mm	-20 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
20.00 mm	19.97 mm	-30 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
25.00 mm	24.97 mm	-30 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$