

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.2 μm	23.3 μm	0.1 μm	± 2.5 μm	±	2.1E-01 μm
51.1 μm	50.6 μm	-0.5 μm	± 2.5 μm	±	2.1E-01 μm
125.1 μm	128 μm	2.9 μm	± 3.8 μm	±	2.1E-01 μm
478.0 μm	482 μm	4.0 μm	± 14.3 μm	±	2.2E-01 μm
952 μm	956 μm	4.0 μm	± 28.6 μm	±	2.5E-01 μm