

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.3 μm	26 μm	2.7 μm	$\pm 1.2 \mu\text{m}$	\pm	6.1E-01 μm
51.6 μm	52 μm	0.4 μm	$\pm 2.5 \mu\text{m}$	\pm	6.1E-01 μm
124.6 μm	126 μm	1.4 μm	$\pm 3.2 \mu\text{m}$	\pm	6.1E-01 μm
250.3 μm	248 μm	-2.3 μm	$\pm 4.5 \mu\text{m}$	\pm	6.1E-01 μm
486.0 μm	486 μm	0.0 μm	$\pm 6.9 \mu\text{m}$	\pm	6.2E-01 μm
963.0 μm	964 μm	1.0 μm	$\pm 12 \mu\text{m}$	\pm	6.3E-01 μm