

Clase de Exactitud (Accuracy Class): Clase I (Class I)

---

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 1	1.000	0.12	± 0.4	± 4.3E-01 mm					
0 a 2	2.000	0.21	± 0.6	± 4.7E-01 mm					
0 a 3	3.000	0.55	± 0.8	± 4.7E-01 mm					
0 a 4	4.000	0.66	± 1	± 4.7E-01 mm					
0 a 5	5.000	0.81	± 1.2	± 4.7E-01 mm					
0 a 6	6.000	0.61	± 1.4	± 3.4E-01 mm					
0 a 7	7.000	0.99	± 1.6	± 3.4E-01 mm					
0 a 8	8.000	0.96	± 1.8	± 3.4E-01 mm					
0 a 9	9.000	1.09	± 2	± 3.4E-01 mm					
0 a 10	10.000	1.13	± 2.2	± 3.4E-01 mm					

---