

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 1	1.000	0.02	± 0.4	± 6.6E-01 mm					
0 a 2	2.000	0.07	± 0.6	± 6.9E-01 mm					
0 a 3	3.000	0.18	± 0.8	± 6.9E-01 mm					
0 a 4	4.000	0.29	± 1	± 6.9E-01 mm					
0 a 5	5.000	0.43	± 1.2	± 6.9E-01 mm					
0 a 6	6.000	0.52	± 1.4	± 6.0E-01 mm					
0 a 7	7.000	0.68	± 1.6	± 6.0E-01 mm					
0 a 8	8.000	0.67	± 1.8	± 6.0E-01 mm					
0 a 9	9.001	0.77	± 2	± 6.0E-01 mm					
0 a 10	10.001	0.92	± 2.2	± 6.0E-01 mm					
0 a 11	11.001	0.68	± 2.4	± 6.0E-01 mm					
0 a 12	12.002	0.54	± 2.6	± 6.0E-01 mm					
0 a 13	13.002	0.84	± 2.8	± 6.0E-01 mm					
0 a 14	14.002	0.97	± 3	± 6.0E-01 mm					
0 a 15	15.002	1.15	± 3.2	± 6.0E-01 mm					
