

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.100	0.35	± 1.2	± 4.3E-01 mm					
0 a 0.5	0.500	0.40	± 1.6	± 4.3E-01 mm					
0 a 1	0.999	1.19	± 2	± 4.3E-01 mm					
0 a 2	2.000	0.44	± 2.8	± 4.7E-01 mm					
0 a 3	3.000	0.55	± 3.6	± 4.7E-01 mm					
0 a 4	4.000	0.67	± 4.4	± 4.7E-01 mm					
0 a 5	5.000	0.77	± 5.2	± 4.7E-01 mm					
0 a 6	6.000	0.78	± 6	± 3.4E-01 mm					
0 a 7	7.000	0.95	± 6.8	± 3.4E-01 mm					
0 a 8	8.000	1.14	± 7.6	± 3.4E-01 mm					
