

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.3 μm	24 μm	0.2 μm	$\pm 1.2 \mu\text{m}$	\pm	5.8E-01 μm
51.6 μm	51 μm	-0.6 μm	$\pm 2.5 \mu\text{m}$	\pm	5.8E-01 μm
124.6 μm	124 μm	-0.6 μm	$\pm 3.2 \mu\text{m}$	\pm	5.8E-01 μm
250.3 μm	251 μm	0.7 μm	$\pm 4.5 \mu\text{m}$	\pm	5.8E-01 μm
486.0 μm	484 μm	-2.0 μm	$\pm 6.9 \mu\text{m}$	\pm	5.8E-01 μm