

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
0 °C	-0.181 °C	0.1 °C	0.281 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.941 °C	15.1 °C	0.159 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.494 °C	30.5 °C	0.006 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.834 %HR	31.9 %HR	2.066 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.261 %HR	51.7 %HR	2.439 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.059 %HR	69.4 %HR	-0.659 %HR	± 3.0 %HR	± 1.9E+00 %HR