

<b>Tipo de Medida</b> ( <i>Measurement Type</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>E.M.P</b> ( <i>Tolerance</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
Exterior / (Exterior)	40.00 mm	40.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	80.00 mm	80.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	120.00 mm	120.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	160.00 mm	160.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	200.00 mm	200.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Interior / (Interior)	20.00 mm	20.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 8.2\text{E}+00 \mu\text{m}$
Interior / (Interior)	180.00 mm	180.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 8.2\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	20.00 mm	20.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	180.00 mm	180.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.1\text{E}+00 \mu\text{m}$