

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.1	0.100	0.03	± 0.2	± 4.3E-01 mm					
0 a 0.5	0.500	0.02	± 0.3	± 4.3E-01 mm					
0 a 1	1.000	0.08	± 0.4	± 4.3E-01 mm					
0 a 2	2.000	0.19	± 0.6	± 4.7E-01 mm					
0 a 3	3.000	0.30	± 0.8	± 4.7E-01 mm					
0 a 4	4.000	0.41	± 1	± 4.7E-01 mm					
0 a 5	5.000	0.54	± 1.2	± 4.7E-01 mm					
0 a 6	6.000	0.37	± 1.4	± 3.4E-01 mm					
0 a 7	7.001	0.34	± 1.6	± 3.4E-01 mm					
0 a 8	8.001	0.43	± 1.8	± 3.4E-01 mm					
