

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.3 μm	23 μm	-0.3 μm	$\pm 1.2 \mu\text{m}$	\pm	5.8E-01 μm
51.6 μm	52 μm	0.4 μm	$\pm 2.5 \mu\text{m}$	\pm	5.8E-01 μm
124.6 μm	124 μm	-0.6 μm	$\pm 3.2 \mu\text{m}$	\pm	5.8E-01 μm
250.3 μm	249 μm	-1.3 μm	$\pm 4.5 \mu\text{m}$	\pm	5.8E-01 μm
486.0 μm	488 μm	2.0 μm	$\pm 6.9 \mu\text{m}$	\pm	5.8E-01 μm
963.0 μm	965 μm	2.0 μm	$\pm 12 \mu\text{m}$	\pm	6.0E-01 μm