

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.1	0.100	0.33	± 0.6	± 4.3E-01 mm					
0 a 0.5	0.500	0.36	± 0.8	± 4.3E-01 mm					
0 a 1	1.000	0.41	± 1	± 4.3E-01 mm					
0 a 2	2.000	0.49	± 1.4	± 4.7E-01 mm					
0 a 3	3.000	0.60	± 1.8	± 4.7E-01 mm					
0 a 4	4.000	0.72	± 2.2	± 4.7E-01 mm					
0 a 5	5.000	0.97	± 2.6	± 4.7E-01 mm					
0 a 6	6.000	0.98	± 3	± 3.4E-01 mm					
0 a 7	7.000	1.19	± 3.4	± 3.4E-01 mm					
0 a 8	8.000	1.42	± 3.8	± 3.4E-01 mm					
