

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
49.80 mmHg	50 mmHg	0.20 mmHg	± 3 mmHg	±	6.6E-01 mmHg
100.20 mmHg	100 mmHg	-0.20 mmHg	± 3 mmHg	±	6.5E-01 mmHg
150.40 mmHg	150 mmHg	-0.40 mmHg	± 3 mmHg	±	6.6E-01 mmHg
199.40 mmHg	200 mmHg	0.60 mmHg	± 3 mmHg	±	6.8E-01 mmHg
249.50 mmHg	250 mmHg	0.50 mmHg	± 3 mmHg	±	7.5E-01 mmHg
298.70 mmHg	300 mmHg	1.30 mmHg	± 3 mmHg	±	7.5E-01 mmHg