

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
19,897.7 lx	19,895 lx	2.7 lx	± 597 lx	± 9.8E+01 lx
1,035.5 lx	1,031 lx	4.5 lx	± 31 lx	± 1.1E+00 lx
340.7 lx	335 lx	5.7 lx	± 10 lx	± 1.8E-01 lx
30.2 lx	28 lx	2.2 lx	± 2 lx	± 1.8E-01 lx