

Clase de Exactitud (Accuracy Class): Clase III (Class III)

| Valor UBP | Valor Patrón | Error | EMP | Incertidumbre | Valor UBP | Valor Patrón | Error | EMP | Incertidumbre |
|-----------|--------------|-------|-------|---------------|-----------|--------------|-------|-----|---------------|
| m | m | mm | mm | (Uncertainty) | m | m | mm | mm | (Uncertainty) |
| 0 a 0.5 | 0.49907 | 0.930 | ± 1.6 | ± 4.3E-01 mm | | | | | |
| 0 a 1 | 0.99913 | 0.870 | ± 2 | ± 4.3E-01 mm | | | | | |
| 0 a 2 | 1.99922 | 0.780 | ± 2.8 | ± 4.8E-01 mm | | | | | |
| 0 a 3 | 2.99946 | 0.540 | ± 3.6 | ± 4.8E-01 mm | | | | | |
| 0 a 3.75 | 3.74977 | 0.230 | ± 4.2 | ± 4.8E-01 mm | | | | | |
| 0 a 4 | 3.99963 | 0.370 | ± 4.4 | ± 4.8E-01 mm | | | | | |
| 0 a 5 | 4.99972 | 0.280 | ± 5.2 | ± 4.8E-01 mm | | | | | |
| 0 a 6 | 5.99962 | 0.380 | ± 6 | ± 3.6E-01 mm | | | | | |
| 0 a 7 | 6.99971 | 0.290 | ± 6.8 | ± 3.6E-01 mm | | | | | |
| 0 a 7.2 | 7.19981 | 0.190 | ± 6.9 | ± 3.6E-01 mm | | | | | |
