

Tipo de Medida (<i>Measurement Type</i>)	Patrón (<i>Standard</i>)	UBP (<i>UT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)
Profundidad / (Depth)	1.00 mm	1.00 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 6.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	1.25 mm	1.26 mm	10 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	1.50 mm	1.51 mm	10 μm	$\pm 30 \mu\text{m}$	$\pm 6.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	2.00 mm	2.01 mm	10 μm	$\pm 30 \mu\text{m}$	$\pm 6.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	5.00 mm	5.01 mm	10 μm	$\pm 30 \mu\text{m}$	$\pm 6.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	10.00 mm	10.02 mm	20 μm	$\pm 30 \mu\text{m}$	$\pm 6.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	15.00 mm	15.02 mm	20 μm	$\pm 30 \mu\text{m}$	$\pm 6.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	20.00 mm	20.02 mm	20 μm	$\pm 30 \mu\text{m}$	$\pm 6.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	25.00 mm	25.02 mm	20 μm	$\pm 30 \mu\text{m}$	$\pm 6.1\text{E}+00 \mu\text{m}$