

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP m	Valor Patrón m	Error mm	EMP mm	Incertidumbre (Uncertainty)	Valor UBP m	Valor Patrón m	Error mm	EMP mm	Incertidumbre (Uncertainty)
0 a 0.1	0.100	0.300	± 0.6	± 4.3E-01 mm					
0 a 0.2	0.200	0.310	± 0.6	± 4.3E-01 mm					
0 a 0.4	0.400	0.280	± 0.7	± 4.3E-01 mm					
0 a 0.6	0.600	0.230	± 0.8	± 4.3E-01 mm					
0 a 0.8	0.800	0.310	± 0.9	± 4.3E-01 mm					
0 a 1	1.000	0.390	± 1	± 4.3E-01 mm					
0 a 1.5	1.500	0.280	± 1.2	± 4.7E-01 mm					
0 a 2	2.000	0.290	± 1.4	± 4.7E-01 mm					
0 a 2.5	2.500	0.410	± 1.6	± 4.7E-01 mm					
0 a 3	3.000	0.520	± 1.8	± 4.7E-01 mm					
0 a 3.5	3.500	0.490	± 2	± 4.7E-01 mm					
0 a 4	4.000	0.490	± 2.2	± 4.7E-01 mm					
0 a 4.5	4.500	0.600	± 2.4	± 4.7E-01 mm					
0 a 5	5.000	0.600	± 2.6	± 4.7E-01 mm					
0 a 5.5	5.500	0.670	± 2.8	± 3.4E-01 mm					
0 a 6	6.000	0.720	± 3	± 3.4E-01 mm					
0 a 6.5	6.500	0.412	± 3.2	± 3.4E-01 mm					
0 a 7	7.000	0.386	± 3.4	± 3.4E-04 mm					
0 a 7.5	7.500	0.325	± 3.6	± 3.4E-04 mm					
0 a 8	7.999	0.565	± 3.8	± 3.4E-04 mm					