

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
8 °C	7.891 °C	8.1 °C	0.209 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.015 °C	20.2 °C	0.185 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.349 °C	32.4 °C	0.051 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.202 %HR	32.6 %HR	2.398 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	52.2 %HR	2.172 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	71.613 %HR	71.1 %HR	-0.513 %HR	± 3.0 %HR	± 1.9E+00 %HR