

| Nominal (<i>Nominal</i>) | Patrón (<i>Standard</i>) | UBP (<i>UUT</i>) | Error (<i>Error</i>) | Exactitud (<i>Accuracy</i>) | Incertidumbre (<i>Uncertainty</i>) |
|-----------------------------------|-----------------------------------|---------------------------|-------------------------------|--------------------------------------|---------------------------------------------|
| 8 °C | 7.891 °C | 7.9 °C | 0.009 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 20 °C | 20.015 °C | 20.1 °C | 0.085 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 32 °C | 32.349 °C | 32.5 °C | 0.151 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 %HR | 30.202 %HR | 32.9 %HR | 2.698 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 50.028 %HR | 52.3 %HR | 2.272 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 72 %HR | 71.613 %HR | 71.7 %HR | 0.087 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |