

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
8 °C	7.891 °C	8.1 °C	0.209 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.015 °C	20.1 °C	0.085 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.349 °C	32.4 °C	0.051 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.202 %HR	32.0 %HR	1.798 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	51.5 %HR	1.472 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	71.613 %HR	71.1 %HR	-0.513 %HR	± 3.0 %HR	± 1.9E+00 %HR