

Certificado No. (Certificate #): 66741  
Fecha de Calibración (Calibration Date): 2026-03-06

Equipo (Instrument): Calibrador Pie de Rey  
Marca (Brand): Starrett

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	0.5000 in	0.4995 in	-0.0005 in	± 0.0012 in	± 3.4E-04 in
Exterior / (Exterior)	1.0000 in	1.0000 in	0.0000 in	± 0.0012 in	± 3.4E-04 in
Exterior / (Exterior)	3.0000 in	3.0000 in	0.0000 in	± 0.0012 in	± 3.4E-04 in
Exterior / (Exterior)	5.0000 in	5.0005 in	0.0005 in	± 0.0012 in	± 3.4E-04 in
Exterior / (Exterior)	6.0000 in	6.0005 in	0.0005 in	± 0.0012 in	± 3.4E-04 in
Interior / (Interior)	1.5000 in	1.5000 in	0.0000 in	± 0.0012 in	± 3.7E-04 in
Interior / (Interior)	5.0000 in	5.0000 in	0.0000 in	± 0.0012 in	± 3.7E-04 in
Profundidad / (Depth)	2.0000 in	2.0000 in	0.0000 in	± 0.0012 in	± 3.3E-04 in
Profundidad / (Depth)	4.0000 in	4.0000 in	0.0000 in	± 0.0012 in	± 3.3E-04 in