

Certificado No. (Certificate #):

66742

Equipo (Instrument): Calibrador Pie de Rey

Fecha de Calibración (Calibration Date):

2026-03-06

Marca (Brand):

Starrett

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	1.0000 in	1.0000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	2.0000 in	2.0000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	3.0000 in	3.0000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	5.0000 in	5.0000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Exterior / (Exterior)	7.5000 in	7.5000 in	0.0000 in	± 0.0012 in	± 3.5E-04 in
Interior / (Interior)	4.0000 in	4.0000 in	0.0000 in	± 0.0012 in	± 3.7E-04 in
Interior / (Interior)	6.0000 in	6.0000 in	0.0000 in	± 0.0012 in	± 3.7E-04 in
Profundidad / (Depth)	3.0000 in	3.0000 in	0.0000 in	± 0.0012 in	± 3.3E-04 in
Profundidad / (Depth)	6.0000 in	6.0000 in	0.0000 in	± 0.0012 in	± 3.3E-04 in