

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.343 °C	30.3 °C	-0.043 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.302 %HR	38 %HR	3.698 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	53 %HR	2.972 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.568 %HR	71 %HR	1.432 %HR	± 5.0 %HR	± 1.9E+00 %HR