

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
0.0233 mm	0.02 mm	-3.30 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
0.1251 mm	0.12 mm	-5.10 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
0.2503 mm	0.25 mm	-0.30 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
0.4860 mm	0.48 mm	-6.00 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
0.9630 mm	0.96 mm	-3.00 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.00 mm	1.00 mm	0.00 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.25 mm	1.25 mm	0.00 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.50 mm	1.50 mm	0.00 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
2.00 mm	2.00 mm	0.00 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
5.00 mm	5.00 mm	0.00 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$