

<b>Tipo de Medida</b> ( <i>Measurement Type</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>E.M.P</b> ( <i>Tolerance</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
Exterior / (Exterior)	45 mm	45.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.6\text{E}+01 \mu\text{m}$
Exterior / (Exterior)	54 mm	54.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.6\text{E}+01 \mu\text{m}$
Exterior / (Exterior)	83 mm	83.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.6\text{E}+01 \mu\text{m}$
Exterior / (Exterior)	73 mm	73.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.6\text{E}+01 \mu\text{m}$
Exterior / (Exterior)	87 mm	87.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.6\text{E}+01 \mu\text{m}$
Exterior / (Exterior)	95 mm	95.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.6\text{E}+01 \mu\text{m}$
Exterior / (Exterior)	126 mm	126.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.6\text{E}+01 \mu\text{m}$
Exterior / (Exterior)	128 mm	128.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.6\text{E}+01 \mu\text{m}$
Exterior / (Exterior)	131 mm	131.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.6\text{E}+01 \mu\text{m}$
Interior / (Interior)	45 mm	45.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.0\text{E}+01 \mu\text{m}$
Interior / (Interior)	131 mm	131.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.0\text{E}+01 \mu\text{m}$
Profundidad / (Depth)	45 mm	45.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.1\text{E}+01 \mu\text{m}$
Profundidad / (Depth)	131 mm	131.00 mm	0 $\mu\text{m}$	$\pm 50 \mu\text{m}$	$\pm 3.1\text{E}+01 \mu\text{m}$