

Certificado No. (Certificate #):

66750

Equipo (Instrument): Reloj Indicador

Fecha de Calibración (Calibration Date):

2026-03-06

Marca (Brand): TECLOCK

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0.050 in	0.050 in	0.000 in	± 0.0014 in	± 5.8E-04 in
0.100 in	0.100 in	0.000 in	± 0.0014 in	± 5.8E-04 in
0.150 in	0.150 in	0.000 in	± 0.0014 in	± 5.9E-04 in
0.200 in	0.200 in	0.000 in	± 0.0014 in	± 5.9E-04 in
0.300 in	0.300 in	0.000 in	± 0.0014 in	± 6.1E-04 in
0.400 in	0.400 in	0.000 in	± 0.0014 in	± 6.4E-04 in
0.500 in	0.500 in	0.000 in	± 0.0014 in	± 6.8E-04 in
0.800 in	0.800 in	0.000 in	± 0.0014 in	± 8.0E-04 in
1.000 in	1.000 in	0.000 in	± 0.0014 in	± 9.1E-04 in