

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
18 °C	18.460 °C	18.1 °C	-0.360 °C	± 0.5 °C	± 4.0E-01 °C
23 °C	23.414 °C	23.2 °C	-0.214 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	28.247 °C	28.5 °C	0.253 °C	± 0.5 °C	± 4.0E-01 °C
45 %HR	44.057 %HR	46.7 %HR	2.643 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	59.578 %HR	60.0 %HR	0.422 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.008 %HR	73.5 %HR	-1.508 %HR	± 3.0 %HR	± 1.9E+00 %HR