

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.146 °C	2.6 °C	0.454 °C	± 1.0 °C	± 4.0E-01 °C
5 °C	4.993 °C	5.3 °C	0.307 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	8.102 °C	8.5 °C	0.398 °C	± 1.0 °C	± 4.0E-01 °C