

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.5	0.49909	0.910	± 1.6	± 4.3E-01 mm					
0 a 1	0.99906	0.940	± 2	± 4.3E-01 mm					
0 a 2	1.99913	0.870	± 2.8	± 4.8E-01 mm					
0 a 3	2.99941	0.590	± 3.6	± 4.8E-01 mm					
0 a 3.75	3.75001	-0.010	± 4.2	± 4.8E-01 mm					
0 a 4	3.99860	1.400	± 4.4	± 4.8E-01 mm					
0 a 5	4.99928	0.720	± 5.2	± 4.8E-01 mm					
0 a 6	5.99907	0.930	± 6	± 3.6E-01 mm					
0 a 7	6.99999	0.010	± 6.8	± 3.6E-01 mm					
0 a 7.2	7.20002	-0.020	± 6.9	± 3.6E-01 mm					
