

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C | 30.062 °C | 30.4 °C | 0.338 °C | ± 1.0 °C | ± 4.0E-01 °C |
| 30 %HR | 30.570 %HR | 32.0 %HR | 1.430 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 50.355 %HR | 51.0 %HR | 0.645 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 70 %HR | 70.059 %HR | 69.0 %HR | -1.059 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |