

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP m	Valor Patrón m	Error mm	EMP mm	Incertidumbre (Uncertainty)	Valor UBP m	Valor Patrón m	Error mm	EMP mm	Incertidumbre (Uncertainty)
0 a 0.05	0.05026	-0.260	± 0.6	± 4.3E-01 mm					
0 a 0.1	0.10025	-0.250	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.15023	-0.230	± 0.6	± 4.3E-01 mm					
0 a 1	1.00047	-0.470	± 1	± 4.3E-01 mm					
0 a 2	2.00020	-0.200	± 1.4	± 4.8E-01 mm					
0 a 3	2.99917	0.830	± 1.8	± 4.8E-01 mm					
0 a 4	3.99932	0.680	± 2.2	± 4.8E-01 mm					
0 a 5	4.99965	0.350	± 2.6	± 4.8E-01 mm					
0 a 6	5.99965	0.350	± 3	± 3.6E-01 mm					
0 a 7	6.99933	0.670	± 3.4	± 3.6E-01 mm					
0 a 8	7.99968	0.320	± 3.8	± 3.6E-01 mm					
0 a 9	8.99913	0.870	± 4.2	± 3.6E-01 mm					
0 a 10	10.00131	-1.310	± 4.6	± 3.6E-01 mm					
0 a 11	11.00007	-0.070	± 5	± 3.6E-01 mm					
0 a 12	12.00037	-0.370	± 5.4	± 3.6E-01 mm					
0 a 13	12.99909	0.910	± 5.8	± 3.6E-01 mm					
0 a 14	13.99950	0.500	± 6.2	± 3.6E-01 mm					
0 a 15	14.99938	0.620	± 6.6	± 3.6E-01 mm					
0 a 16	15.99973	0.270	± 7	± 3.6E-01 mm					
0 a 17	16.99813	1.870	± 7.4	± 3.6E-01 mm					
0 a 18	17.99896	1.040	± 7.8	± 3.6E-01 mm					
0 a 19	18.99935	0.650	± 8.2	± 3.6E-01 mm					
0 a 20	20.00016	-0.160	± 8.6	± 3.6E-01 mm					