

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.4907 µl	0.5 µl	0.0093 µl	± 0.050 µl	±	5.1E-02 µl
4.9975 µl	5 µl	0.0025 µl	± 0.125 µl	±	5.1E-02 µl
10.010 µl	10 µl	-0.0101 µl	± 0.120 µl	±	5.1E-02 µl