

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
2 °C	2.095 °C	2.0 °C	-0.095 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.212 °C	8.1 °C	-0.112 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.413 °C	30.3 °C	-0.113 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.139 %HR	42.1 %HR	2.961 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.711 %HR	66.9 %HR	2.189 %HR	± 3.0 %HR	± 1.9E+00 %HR