

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.095 °C	2.3 °C	0.205 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.212 °C	8.3 °C	0.088 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.413 °C	30.6 °C	0.187 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.139 %HR	41.5 %HR	2.361 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.711 %HR	65.0 %HR	0.289 %HR	± 3.0 %HR	± 1.9E+00 %HR