

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.095 °C	2.1 °C	0.005 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.212 °C	8.2 °C	-0.012 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.413 °C	30.3 °C	-0.113 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.139 %HR	42.0 %HR	2.861 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.711 %HR	66.3 %HR	1.589 %HR	± 3.0 %HR	± 1.9E+00 %HR