

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.095 °C	1.9 °C	-0.195 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.212 °C	8.1 °C	-0.112 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.413 °C	30.4 °C	-0.013 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	39.139 %HR	42.1 %HR	2.961 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.711 %HR	64.4 %HR	-0.311 %HR	± 3.0 %HR	± 1.9E+00 %HR