

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.494 °C	30.3 °C	-0.194 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.834 %HR	31.9 %HR	2.066 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.261 %HR	51.4 %HR	2.139 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.059 %HR	69.9 %HR	-0.159 %HR	± 3.0 %HR	± 1.9E+00 %HR