

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.494 °C	30.5 °C	0.006 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.834 %HR	31.5 %HR	1.666 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.261 %HR	51.7 %HR	2.439 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.059 %HR	70.1 %HR	0.041 %HR	± 3.0 %HR	± 1.9E+00 %HR