

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C | 30.494 °C | 30.4 °C | -0.094 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 %HR | 29.834 %HR | 32.6 %HR | 2.766 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 49.261 %HR | 52.0 %HR | 2.739 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 70 %HR | 70.059 %HR | 70.6 %HR | 0.541 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |