

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>	
0.0516 mm	0.051 mm	-0.60 $\mu\text{m}$	$\pm 3 \mu\text{m}$	$\pm$	5.8E-01 $\mu\text{m}$
0.0233 mm	0.024 mm	0.70 $\mu\text{m}$	$\pm 3 \mu\text{m}$	$\pm$	5.8E-01 $\mu\text{m}$
0.1251 mm	0.125 mm	-0.10 $\mu\text{m}$	$\pm 3 \mu\text{m}$	$\pm$	5.8E-01 $\mu\text{m}$
0.2503 mm	0.250 mm	-0.30 $\mu\text{m}$	$\pm 3 \mu\text{m}$	$\pm$	5.8E-01 $\mu\text{m}$
0.4860 mm	0.486 mm	0.00 $\mu\text{m}$	$\pm 3 \mu\text{m}$	$\pm$	5.8E-01 $\mu\text{m}$
0.9630 mm	0.964 mm	1.00 $\mu\text{m}$	$\pm 3 \mu\text{m}$	$\pm$	5.8E-01 $\mu\text{m}$