

Patrón (Standard)		UBP (UIT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
10 Hz	0.800 g-pk	0.83 g-pk	0.03 g-pk	± 0.04 g-pk	±	1.7E-02 g-pk
30 Hz	1.000 g-pk	1.04 g-pk	0.04 g-pk	± 0.05 g-pk	±	2.1E-02 g-pk
50 Hz	1.000 g-pk	0.98 g-pk	-0.02 g-pk	± 0.05 g-pk	±	2.1E-02 g-pk
100 Hz	1.000 g-pk	0.98 g-pk	-0.02 g-pk	± 0.05 g-pk	±	2.1E-02 g-pk
500 Hz	1.000 g-pk	1.03 g-pk	0.03 g-pk	± 0.05 g-pk	±	2.1E-02 g-pk
1000 Hz	1.000 g-pk	0.97 g-pk	-0.03 g-pk	± 0.05 g-pk	±	2.2E-02 g-pk