

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
35 °C	35.267 °C	35.4 °C	0.133 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.509 %HR	29.7 %HR	-0.809 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.752 %HR	46.0 %HR	-3.752 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.110 %HR	66.1 %HR	-4.010 %HR	± 5.0 %HR	± 1.9E+00 %HR