

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
35 °C	35.267 °C	35.2 °C	-0.067 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.509 %HR	30.8 %HR	0.291 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.752 %HR	49.9 %HR	0.148 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.110 %HR	73.2 %HR	3.090 %HR	± 5.0 %HR	± 1.9E+00 %HR